



COMMITTEE DRAFT ISO/CD 10596

Date 2008-02-01	Reference number <b>ISO/TC 8 / SC 6 N 189</b>
Supersedes document	

**WARNING:** This document is not an International Standard. It is distributed for review and comment. It is subject to change without notice and may not be referred to as an International Standard.

**ISO/TC 8 / SC 6**

Title  
**Navigation**

Secretariat **JISC/JSTRA**

Circulated to P- and O-members, and to technical committees and organizations in liaison for:

discussion at \_\_\_\_\_ on \_\_\_\_\_  
[venue/date of meeting]

comments by **2008-05-02**  
[date]

approval for registration as a DIS in accordance with 2.5.6 of part 1 of the ISO/IEC Directives, by

**2008-05-02**  
[date]

(P-members vote only: ballot form attached)

**P-members of the technical committee or subcommittee concerned have an obligation to vote.**

English title

**Ships and marine technology – Marine wind vane and anemometers**

French title

Reference language version:  English  French  Russian

Introductory note

The SC 6 Secretariat informed you in its documents (doc. TC 8/SC 6 N 189) that ISO/CD 10596 would be developed by Expert Group of ISO 10596. (Project leader : Dr. Tatsuo Hanafusa, Japan)

This draft is circulated as ISO/CD 10596 in order to carry on the following two.

- collecting SC 6 member's comments with regard to the contents of the draft.
- voting for agree or against "making the next stage of draft deliberation into the stage of having ISO/DIS 10596."



ISO TC 8/SC 6 N **189**

Date: 2008-02-01

**ISO/CD 10596**

ISO TC 8/SC 6/WG 13

Secretariat: JISC/JSTRA

## **Ships and marine technology — Marine wind vane and anemometers**

### **Warning**

This document is not an ISO International Standard. It is distributed for review and comment. It is subject to change without notice and may not be referred to as an International Standard.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Document type: International Standard  
Document subtype:  
Document stage: (30) Committee  
Document language: E

### Copyright notice

This ISO document is a working draft or committee draft and is copyright-protected by ISO. While the reproduction of working drafts or committee drafts in any form for use by participants in the ISO standards development process is permitted without prior permission from ISO, neither this document nor any extract from it may be reproduced, stored or transmitted in any form for any other purpose without prior written permission from ISO.

Requests for permission to reproduce this document for the purpose of selling it should be addressed as shown below or to ISO's member body in the country of the requester:

[Indicate the full address, telephone number, fax number, telex number, and electronic mail address, as appropriate, of the Copyright Manager of the ISO member body responsible for the secretariat of the TC or SC within the framework of which the working document has been prepared.]

Reproduction for sales purposes may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

**Contents**

Page

<b>Foreword</b> .....	<b>iv</b>
<b>1</b> <b>Scope</b> .....	<b>1</b>
<b>2</b> <b>Normative references</b> .....	<b>1</b>
<b>3</b> <b>Terms and definitions</b> .....	<b>1</b>
<b>4</b> <b>Type</b> .....	<b>2</b>
<b>5</b> <b>Composition</b> .....	<b>2</b>
<b>6</b> <b>Functionality</b> .....	<b>2</b>
<b>7</b> <b>Performance and accuracy</b> .....	<b>2</b>
<b>8</b> <b>Test</b> .....	<b>4</b>
<b>9</b> <b>Electromagnetic compatibility test</b> .....	<b>7</b>
<b>10</b> <b>Marking</b> .....	<b>7</b>
<b>11</b> <b>Installation</b> .....	<b>8</b>
<b>12</b> <b>User's manual</b> .....	<b>8</b>
<b>Bibliography</b> .....	<b>9</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 10596 was prepared by Technical Committee ISO/TC 8, *Ships and marine technology*, Subcommittee SC 6, *Navigation*.

# Ships and marine technology — Marine wind vane and anemometers

## 1 Scope

This International Standard specifies the type, structure, function, performance, and testing method of marine wind vane and anemometers (hereafter referred to as “wind vane/anemometer”) to be installed on a ship for measurement and indication of meteorological data (wind direction and velocity) at sea, as recommended by Regulation 5 of Chapter V of the international convention for the Safety of Life at Sea (SOLAS), 1974, as amended.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 61162-1, *Maritime navigation and radiocommunication equipment and systems - Digital interfaces - Part 1: Single talker and multiple listeners*

IEC 60945, *Marine navigational and radiocommunication equipment and systems - General requirements - Methods of testing and required test results.*

## 3 Terms and definitions

For the purposes of this International Standard, the following definitions apply.

### 3.1

#### **wind speed**

magnitude of straight-line moving distance of airflow per unit time

### 3.2

#### **wind direction**

wind direction is the direction from where the wind is blowing from.

### 3.3

#### **wind speed measurement range**

range of measurable wind speed within the accuracy specified in this standard

### 3.4

#### **operating temperature range**

range of airflow temperatures in which wind speed can be measured within the accuracy specified in this standard

### 3.5

#### **distance constant**

The distance the air flows past a rotating anemometer during the time it takes the propeller to reach  $(1-1/e)$  or 63% of the equilibrium speed after a step increase change in air speed.

### 3.6

#### **index error**

the amount by which a wind vane/anemometer exceeds or falls short of the true value

## 4 Type

The types of wind vanes and anemometers shall be classified as follows in accordance with methods of measuring wind direction and wind speed.

### 4.1 Using wind force on a rotor

- a) A rotation anemometer whose axis of rotation is horizontal. The instrument has either flat vanes or helicoidally vanes.
- b) A rotation anemometer whose axis of rotation is vertical. Cup anemometers usually consist of three or four hemispherical or conical cups mounted with their diametrical planes vertical and distributed symmetrically about the axis of rotation.

### 4.2 Using ultrasonic waves

An anemometer which measures ultra sonic wave propagation in the air.

## 5 Composition

A wind vane/anemometer is composed of the wind vane/anemometer sensor (hereafter simply referred to as "sensor"), display, etc. The sensor shall have measurement functions for wind direction and velocity, and the display shall be capable of indicating the measured wind direction and velocity.

## 6 Functionality

A wind vane/anemometer shall have the following functions.

**6.1** The sensor shall have measurement functions for wind direction and velocity, whose range and accuracy are specified 7.1 and 7 respectively, and the display shall be capable of indicating the measured wind direction and velocity.

**6.2** The wind vane/anemometer shall be capable of outputting analog or digital signals, which shall be able to be distributed to the bridge and other necessary locations. Where digital signals are used, at least one of them shall satisfy IEC 61162-1.

## 7 Performance and accuracy

The performance and accuracy required for a wind vane/anemometer shall be as indicated below.

### 7.1 Measurement range and minimum measurement unit

The measurement ranges and minimum measurement units for a wind vane/anemometer shall be as indicated in Table 1.

Table 1 — Measurement ranges and minimum measurement units for a wind vane/anemometer

	Measurement range	Minimum measurement unit
Wind speed	2 m/s to 60 m/s or more	0,5 m/s or less
Wind direction	0° to 359°	10° or less

## 7.2 Measurement accuracy

The measurement accuracy for a wind vane/anemometer shall satisfy the values indicated in Table 2 when inspections specified in 8.1 are conducted.

Table 2 — Accuracy of wind vane/anemometer

Type of wind vane/anemometer	Wind direction accuracy	Wind speed accuracy
Windmill wind vane/anemometer	Tolerance of index error at each measurement point shall be $\pm 5^\circ$	When the measurement velocity is less than 10 m/s, tolerance for index error shall be $\pm 0,5$ m/s for each measurement.  When the measurement velocity exceeds 10 m/s, tolerance for index error shall be $\pm 5\%$ for each measurement.
Cup anemometer		
Ultrasonic wind vane / anemometer		

## 7.3 Starting wind speed (limited to types using wind force on a rotor)

A wind vane/anemometer shall satisfy the following requirements when inspected as specified in 8.2.

- a) At the minimum level of the wind speed measurement range, the wind receiving part (propeller (impeller) or cup) of an anemometer shall start and maintain rotation from any position.
- b) At the minimum level of the wind speed measurement range, the wind-receiving part (blades in windmill type, and tails in cup types) of a wind vane shall remain parallel to the airflow.

## 7.4 Sampling frequency (limited to types using ultrasonic waves)

For an ultrasonic wind vane/anemometer, the sampling frequency shall be four or more times per second.

## 7.5 Distance constant

A wind vane/anemometer shall satisfy the value indicated in Table 3 when inspections specified in 8.3 are conducted.

Table 3 — Distance constant for wind vane/anemometer

Type of wind vane/anemometer	Distance constant
Windmill wind vane/anemometer	9 m or less
Cup anemometer	

## 8 Test

A wind vane/anemometer shall be tested as follows.

### 8.1 Test for measurement accuracy

Test for measurement accuracy of a wind vane/anemometer shall be conducted as follows.

#### 8.1.1 Wind vane

a) Wind vane using wind force on a rotor

- 1) Set a sensor (wind vane in the case of a cup type) vertically at the centre of the horizontally placed wind direction inspection board specified in 8.5.3.
- 2) Align the base scale ( $0^\circ$ ) of the wind direction inspection board correctly with the mark on the sensor indicating the direction of the bow.
- 3) Rotate the sensor clockwise or counter-clockwise as indicated, set the value of the indicator correctly in the given direction, and read the scale of the wind direction inspection board at that moment to determine the index error. The inspection points shall be at  $0^\circ$ ,  $90^\circ$ ,  $180^\circ$ ,  $360^\circ$  ( $0^\circ$ ).
- 4) The value at each measurement point shall satisfy the accuracy specified in Table 2.

b) Ultrasonic type

- 1) Set a wind direction inspection board that can rotate  $360^\circ$  on the inspection table of a wind tunnel meeting the requirements in 8.5.1, and set the sensor on the top.
- 2) Align the direction of the wind tunnel wind speed axis and the mark on the sensor indicating the direction of the bow.
- 3) Reduce the wind speed of the wind tunnel to below 2 m/s, rotate the sensor and change the value of the indicator by  $90^\circ$ , take the reading on the wind direction inspection board, and determine the index error. This inspection shall be conducted both clockwise and counter-clockwise.
- 4) The value at each measurement point shall satisfy the accuracy specified in Table 2.

#### 8.1.2 Anemometer

Anemometer inspections shall be conducted as follows using the wind tunnel specified in 8.5.1.

- a) Wind velocities to be measured shall be the lower limit of the measurement range, 5, 10, and 30 m/s, and the upper limit of the measurement range. However, the wind speed of the wind tunnel does not need to reflect these figures exactly; inspections can be conducted using approximate values.

- b) Indications of the wind speed of the wind tunnel and the anemometer shall be measured under stable wind speed in increments of 0,1 m/s. Tolerance below 10 m/s is  $\pm 0,5$  m/s and above 10 m/s is  $V \times 5\%$ .
- c) The value at each measurement point shall satisfy the accuracy specified in Table 2.

## 8.2 Inspection on starting wind speed (limited to types using wind force on a rotor)

An inspection on start wind speed of a wind vane/anemometer shall be conducted as follows, at less than the lower limit of the wind speed measurement range, to confirm that the following conditions are satisfied.

### 8.2.1 Wind vane

- a) Use a wind tunnel meeting the requirements in 8.5.1 and set the wind speed below the lower limit of the measurement range of the anemometer.
- b) Hold the wind-receiving part of the wind vane at a right angle to the airflow for a period, and then release. Confirm that this part sets itself and remains stationary in parallel with the airflow. This inspection shall be conducted clockwise and counter-clockwise.
- c) The wind-receiving part of the wind vane shall remain stationary in parallel with the airflow whether facing clockwise or counter-clockwise.

### 8.2.2 Anemometer

- a) Use a wind tunnel meeting the requirements in 8.5.1 and set the wind speed below the lower limit of the anemometer.
- b) Hold the wind-receiving part of the anemometer at a given position, and then release. Confirm that it resumes rotation from several initial positions.
- c) The wind-receiving part of the anemometer (propeller (impeller) or cups) shall remain rotating from whichever position it starts rotation.

## 8.3 Distance constant inspection (limited to types using wind force on a rotor)

Distance constant inspections shall be conducted as follows, using the wind tunnel specified in 8.5.1.

- a) The velocity,  $V$ , of the wind tunnel specified in 8.5.1 shall be set at [approximately 10 m/s].
- b) Hold the wind-receiving part of the anemometer in a stationary position and then release it. Record the elapsed time in seconds to reach  $0,63 V$  (m/s). This is the time constant of the anemometer.
- c) The distance constant can be calculated by multiplying the measured wind speed  $V$  (m/s) by the time constant  $S$  (s).
- d) The calculated distance constant shall satisfy the value specified in Table 3.

## 8.4 Test methods in which a wind tunnel is not used

For an anemometer certified by a public or a quasi-public body and whose wind speed as well as velocity of the windmill or wind-cup rotation are expressly verified, the wind speed test may be conducted on rotational testing equipment meeting the requirements in 8.5.2.

As for the start wind speed and wind direction, if the relevant starting torque of the equipment has been clearly verified, measurements on the starting torque can be utilized as replacements of the above mentioned inspection.

## 8.5 Test facilities

Requirements for test facilities used for the inspection of a wind vane/anemometer shall be as follows.

### 8.5.1 Wind tunnel

The wind tunnel utilized for inspection of a wind vane/anemometer shall generate sufficiently regulated airflow, enable various inspections to be conducted without affecting the sensor of the wind vane/anemometer, remain traceable to national standards, and shall be appropriately maintained.

However, the wind tunnel may include equipment that generates relative airflows in a stationary atmosphere by moving an object (e.g., running vehicles).

### 8.5.2 Rotary testing equipment

The equipment shall be capable of providing stable rotation to the rotating axis of an anemometer and shall allow reading of rotational velocity.

### 8.5.3 Wind direction inspection board

A wind direction inspection board used to determine index error of a wind vane shall be as follows.

- a) For the inspection board designed for the wind vane using the force on the rotor, the centre of the rotation axis of the wind vane shall be able to be set vertically on the centre of the wind direction inspection board, and the rotation angles of the wind vane shall be measurable in increments of 1°.
- b) The wind direction inspection board shall bear a mark indicating the direction of the bow, and shall be scaled from 0° to 359° from 0° clockwise.
- c) For a wind direction inspection board using ultrasonic waves, when setting the wind direction inspection board on the wind tunnel inspection table, the rotation angles of the wind direction inspection board shall be readable in increments of 1°.
- d) For a wind direction inspection board used to inspect a wind vane using wind force on a rotor, the rotation axis of the wind vane shall be fixed vertically to the centre. In addition, the wind direction inspection board shall be functionally capable of aligning the mark indicating the direction of the bow with that of the wind vane in the direction of its bow, and of allowing measurements to be readable in increments of 1°.

For a wind direction inspection board used to inspect a wind vane using the force on the rotor, the rotation axis of the wind vane shall be fixed vertically to the centre. In addition, the wind direction inspection board shall be functionally capable of aligning the mark indicating the direction of the bow with that of the wind vane. Last, the board shall allow measurements to be readable in increments of 1°.

## 8.6 Environmental conditions and inspections

### 8.6.1 Use conditions

Wind vane/anemometer should be used under the following conditions.

- a) Operating temperature range
  - Equipment installed outdoors - 25 °C to + 55 °C
  - Equipment installed indoors - 15 °C to + 55 °C
- b) Operating time

— Continuous

c) Operating power range

The operating level can be set from 100 V to 240 V for AC voltage, or from 12 V to 24 V for DC voltage.

The tolerance level for the set figures shall satisfy the requirements of IEC 60945.

### 8.6.2 Environmental testing

Environmental tests for a wind vane/anemometer shall be conducted as follows.

- a) A wind vane/anemometer shall satisfy the testing regulations governing environmental conditions for exposed equipment, as specified in Clause 8 of IEC 60945. However, the “rainwater and spray test” for the wind vane/anemometer shall be conducted as follows.
  - b) Rainwater and spray test
    - 1) The rainwater and spray test shall be conducted using a testing device in which the inner diameter of the nozzle is 6,3 mm and the flow rate is 12,5 l/min. Water shall be sprayed at an angle of 60° to the upward vertical direction while the wind vane/anemometer is set in the position at which observations are ordinarily performed.
    - 2) After the test, the results shall satisfy the test results specified in 8.8.3 of IEC 60945, and shall indicate no damage or effects on performance of the wind vane/anemometer.

## 9 Electromagnetic compatibility test

A wind vane/anemometer, with electromagnetic interference and immunity, shall comply with clause 9 and 10 of IEC 60945.

## 10 Marking

### 10.1 Identification plate

Each wind vane/anemometer shall be equipped with an identification plate showing the following in an easily viewable location.

- Name and type of wind vane/anemometer
- Measurement range
- Production number
- Manufacturing date or code
- Manufacturer or code

### 10.2 Azimuth marker

A marker to align the base position of the wind direction (0°) with the direction of the bow shall be placed at an appropriate place on the sensor (on the vane for cup type) of the wind vane/anemometer.

## 11 Installation

Efforts shall be made to install a wind vane/anemometer so as to minimize the effects of the vessel's structure, electromagnetic waves emitted from vessel radars, etc.

## 12 User's manual

The manufacturer shall provide a User's Manual with each wind vane/anemometer. The Manual shall include the following:

- a) An overview of the equipment and its performance
- b) Sufficient information to enable a competent crew member to safely operate, read and maintain the equipment
- c) Adequate instructions, including diagrams, to enable safe installation of the equipment.

## Bibliography

- [1] ISO 17731-1, Meteorology – Wind measurements – Part 1: Wind tunnel test methods for rotating anemometer performance.